

Alaska cont'd

AK102

Method for determination of aromatic and aliphatic hydrocarbons in diesel range organics.

DRO Mix (Tennessee/Mississippi) (16 components)

<i>n</i> -decane (C10)	<i>n</i> -octadecane (C18)
<i>n</i> -undecane (C11)	<i>n</i> -nonadecane (C19)
<i>n</i> -dodecane (C12)	<i>n</i> -eicosane (C20)
<i>n</i> -tridecane (C13)	<i>n</i> -heneicosane (C21)
<i>n</i> -tetradecane (C14)	<i>n</i> -docosane (C22)
<i>n</i> -pentadecane (C15)	<i>n</i> -tricosane (C23)
<i>n</i> -hexadecane (C16)	<i>n</i> -tetracosane (C24)
<i>n</i> -heptadecane (C17)	<i>n</i> -pentacosane (C25)

1,000 μ g/mL each in methylene chloride, 1mL/ampul
cat. # 31214 (ea.)

Kerosene Composite Standard

5,000 μ g/mL in methylene chloride, 1mL/ampul
cat. # 31094 (ea.)

50,000 μ g/mL in methylene chloride, 1mL/ampul
cat. # 31256 (ea.)

50,000 μ g/mL in methylene chloride, 5mL/ampul
cat. # 31257 (ea.)

Diesel Fuel #2 Composite Standard

5,000 μ g/mL in methylene chloride, 1mL/ampul
cat. # 31093 (ea.)

50,000 μ g/mL in methylene chloride, 1mL/ampul
cat. # 31258 (ea.)

50,000 μ g/mL in methylene chloride, 5mL/ampul
cat. # 31259 (ea.)

o-Terphenyl

2,000 μ g/mL in acetone, 1mL/ampul
cat. # 31066 (ea.)

10,000 μ g/mL in methylene chloride, 1mL/ampul
cat. # 31097 (ea.)

5- α -androstane

2,000 μ g/mL in methylene chloride, 1mL/ampul
cat. # 31065 (ea.)

Alaska cont'd

AK103

Method for determination of aromatic and aliphatic hydrocarbons in residual range organics.

Residual Range Calibration Standard (RCS)

SAE30 motor oil:SAE40 motor oil (1:1)
50,000 μ g/mL in methylene chloride, 1mL/ampul
cat. # 31817 (ea.)

Residual Range Calibration Verification Standard (CVS)

SAE30 motor oil:SAE40 motor oil (1:1)
25,000 μ g/mL in methylene chloride, 1mL/ampul
cat. # 31818 (ea.)

Motor Oil Composite Standard

50,000 μ g/mL in methylene chloride, 1mL/ampul
cat. # 31464 (ea.)

Fuel Oil #6 Standard

This fuel, sometimes called bunker C or residual, is a black viscous oil. Applications in which it may be used require the ability to preheat the fuel prior to pumping and burning.

5,000 μ g/mL in methylene chloride, 1mL/ampul
cat. # 31218 (ea.)

50,000 μ g/mL in methylene chloride, 1mL/ampul
cat. # 31248 (ea.)

50,000 μ g/mL in methylene chloride, 5mL/ampul
cat. # 31249 (ea.)

n-Triacontane-d62

500 μ g/mL in methylene chloride, 1mL/ampul
cat. # 31816 (ea.)

Surrogate Standard Mixture

squalane tetrahydronaphthol
o-terphenyl
1,000 μ g/mL each in methylene chloride, 1mL/ampul
cat. # 31638 (ea.)

free literature

Alaska UST Monitoring

Download your free copy from
www.restek.com.

Fast Facts

lit. cat. # 59503

free data

Available on Our Website: Lot Certificates, Data Packs, and MSDSs

For complete information detailing manufacturing and testing for Restek inventoried reference standards, visit our website at www.restek.com. To view lot certificates and/or an MSDS, enter the catalog number of the product in the Search feature. For a free data pack (Adobe® PDF file), enter the catalog number and lot number of the product.